

# Temperature dry-well calibrator Model CTD9100-ZERO

WIKA data sheet CT 41.30



for further approvals  
see page 3

## Applications

- Simple calibration of thermometers at 0 °C (32 °F)
- Tempering of thermocouple cold junctions
- Test benches and calibration laboratories

## Special features

- Low-cost
- Low weight and compact design
- Simple operation



Temperature dry-well calibrator model CTD9100-ZERO

## Description

### Versatile in application

Nowadays, fast and simple testing of thermometers is a “must” when it comes to the operational safety of machines and plants. The portable calibrators of the CTD9100 family are particularly suited for on-site calibrations and extremely user-friendly. Due to their compact design and their low weight, the instruments can be taken and used almost anywhere.

### Special application

Regular monitoring of temperature probes helps to reduce failures and production downtimes. These temperature dry-well calibrators not only calibrate, but they can also temper. In a simple way, a lasting cold junction temperature can be generated and thus the calibration of thermocouples is simplified. The complex method of water fixed-point cells and the risk of freezing and damage in transport are avoided.

The freezing-point calibrator can deliver not only the zero point in °C, but also, through the active cooling, it can deliver other test temperatures.

### Features

This instrument concept combines a stable temperature source with a precise Pt100 temperature measurement, which enables industrial temperature probes to be calibrated even more efficiently and in shorter times. Due to its design and the control, an even temperature distribution is achieved within the block. On this basis, the temperature can be taken as homogenous and not as distributed over the seven test bores. This homogenous temperature distribution reduces the influences on measurement uncertainty.

## Easy to use

The CTD9100-ZERO represents the current state of the art with respect to measurement accuracy, functionality and the safe operation of the instrument. The operating position is defined as upright, since this achieves the optimum temperature distribution.

The calibration temperature, adjusted simply using two keys on the controller, can be very quickly controlled.

The set temperature of the heating block is displayed on a large, 4-digit, high-contrast LED display. Thus reading errors are virtually eliminated.

For easy reading of the temperature, the display is not only large, but, in addition, is also angled by 35°, easy to use.

## Specifications

CTD9100-ZERO	
<b>Indication</b>	
Display	4-digit LED display
Temperature range	-10 ... 0 ... +100 °C [14 ... 32 ... 212 °F]
Accuracy <sup>1)</sup>	0.05 K at 0 °C [32 °F] otherwise 0.1 K
Stability <sup>2)</sup>	< 0.05 K
Display resolution	0.1 °C
<b>Temperature distribution</b>	
Axial homogeneity <sup>3)</sup>	dependent on temperature, temperature probes and their quantity
<b>Temperature control</b>	
Heating time	15 min from -10 °C to +100 °C (from 14 °F to 212 °F)
Cooling time	5 min from 23 °C to 0 °C (from 73 °F to 14 °F) 10 min from 100 °C to 0 °C (from 212 °F to 14 °F)
Stabilisation time <sup>4)</sup>	dependent on temperature and temperature probe
<b>Insert</b>	
Immersion depth	150 mm [5.91 in]
Test bores	7 bores with Ø 6.5 mm [0.26 in]
<b>Voltage supply</b>	
Power supply <sup>5)</sup>	AC 100 ... 240 V, 50/60 Hz
Power consumption	225 VA
Fuse	3.15 A slow blow fuse (at AC 250 V)
Power cord	AC 230 V; for Europe
<b>Communication</b>	
Interface	RS-485
<b>Case</b>	
Dimensions (W x D x H)	160 x 230 x 320+50 mm [6.3 x 9.1 x 12.6+2 in]
Weight	7 kg [15.5 lbs]

1) Is defined as the measuring deviation between the measured value and the reference value.

2) Maximum temperature difference at a stable temperature over 30 minutes.

3) Maximum temperature difference at 40 mm [1.57 in] above the bottom.

4) Time before reaching a stable value.

5) AC 115 V power supply must be specified on the order, otherwise an AC 230 V one will be delivered.

The measurement uncertainty is defined as the total measurement uncertainty ( $k = 2$ ), which contains the following shares: accuracy, measurement uncertainty of reference, stability and homogeneity.

## Approvals

Logo	Description	Country
CE	<b>EU declaration of conformity</b> <ul style="list-style-type: none"> <li>■ EMC directive</li> <li>■ Low voltage directive</li> <li>■ RoHS directive</li> </ul>	European Union
GOST	<b>GOST (option)</b> Metrology, measurement technology	Russia
KazInMetr	<b>KazInMetr (option)</b> Metrology, measurement technology	Kazakhstan
-	<b>MTSCHS (option)</b> Permission for commissioning	Kazakhstan
BelGIM	<b>BelGIM (option)</b> Metrology, measurement technology	Belarus

## Certificates


Certificate	
<b>Calibration</b>	Standard: 3.1 calibration certificate per DIN EN 10204 Option: DKD/DAkkS calibration certificate
<b>Recommended recalibration interval</b>	1 year (dependent on conditions of use)

Approvals and certificates, see website

## Display and control panel

- Frequently used set points can be stored in four memory locations.
- The U key is used to retrieve stored set temperatures.
- The arrow keys are used to change the set temperatures.
- The P key is used to confirm the changes.

- ① Temperature display
- ② Recall key
- ③ Increase key
- ④ Decrease key
- ⑤ Programming key



## Accessories

Accessories		Order code
Description		CTX-A-K6
	<b>Power cord</b> for EU	-EU-
	for Switzerland	-CH-
	for UK	-UK-
	for USA/Canada	-US-
	<b>Calibration software</b>	-CS-
<b>Interface converter RS-485 to USB 2.0</b>		-RC-
<b>Ordering information for your enquiry:</b>		
		<b>1. Order code: CTX-A-K6</b> <b>2. Option:</b>
		↓ [   ]

## Scope of delivery

- Temperature dry-well calibrator model CTD9100-ZERO
- Power cord 1.5 m [5 ft] with safety plug
- Operating instructions
- 3.1 calibration certificate per DIN EN 10204

## Options

- Display in Fahrenheit °F
- DKD/DAkkS calibration certificate only at zero point

## Ordering information

Model / Unit / Calibration / Transport case / Power cord / Further approvals / Additional ordering information

© 05/2008 WIKA Alexander Wiegand SE & Co. KG, all rights reserved.  
The specifications given in this document represent the state of engineering at the time of publishing.  
We reserve the right to make modifications to the specifications and materials.



**ICS Schneider Messtechnik GmbH**  
Briesestrasse 59  
D-16562 Hohen Neuendorf / OT Bergfelde  
Tel.: +49 3303 5040-66  
Fax: +49 3303 5040-68  
E-Mail: info@ics-schneider.de



**WIK A Alexander Wiegand SE & Co. KG**  
Alexander-Wiegand-Straße 30  
63911 Klingenberg/Germany  
Tel. +49 9372 132-0  
Fax +49 9372 132-406  
info@wika.de  
www.wika.de